

To access calibration reports, click on the "CRS" tab at www.securityscale.com



Sample Calibration Report:

1-800-572-2140 **Security Scale Service, Inc.** A2LA Cert No. 1820.01
 Bristol, Charlottesville, Harrisonburg, Lynchburg, Richmond, Roanoke, VA
 P.O. Box 2206 Roanoke, VA 24012-2207

ISO / IEC 17025 Calibration Report

"THE RIGHT WEIGH" Certificate Number: 678900123050609

Customer ABC PRODUCTION - ROANOKE	Services During 06/07/2008	Service Status (In Service) Out of Service
Address 102 PRODUCTION LANE ROANOKE, VA 24012	Next Service Due 05/2009	Weight Class F
City, State, Zip ROANOKE, VA 24012	Manufacturer FAIRBANKS	Model # 413132
Location PRODUCTION	Capacity/Resolution 0-500 lb X 1 lb	Serial 0123
Overall Result (P) F A (P: Pass, F: Fail, A: Adjust)	Temp 75 deg. F	Humidity 55%
	Other	Seal Applied Yes No [N/A]

EQUIPMENT CONDITIONS

At Zero	Clean	Dirty	Good	Fair	Poor
---------	-------	-------	------	------	------

Check Point	As Found Weight	Error	As Left Weight	Final Error	Tolerance
50 lb	50 lb	0 lb	50 lb	0 lb	50-50 lb
100 lb	100 lb	0 lb	100 lb	0 lb	100-100 lb
200 lb	200 lb	0 lb	200 lb	0 lb	200-200 lb
500 lb	500 lb	0 lb	500 lb	0 lb	500-500 lb

Tolerance (As Found)	Tolerance (As Left)	Method Used	Data Transfer Checked
[In] Out	[In] Out	SSS SOP [HB44] Customers	Yes No [N/A]

Floor Scale		Bench Balance	
1	2	1	
3	4	4	2
		3	

As Found		As Left	
Reading	Error	Reading	Error
1: 100 lb	0 lb	1: 100 lb	0 lb
2: 100 lb	0 lb	2: 100 lb	0 lb
3: 100 lb	0 lb	3: 100 lb	0 lb
4: 100 lb	0 lb	4: 100 lb	0 lb

Shift Test Results: (P) F A N/A
Shift Weight: 100lb (P: Pass, F: Fail, A: Adjust, N/A: Not Applicable)

Comments/Notes:
 Technician: dbrayn
 Weight Series/Kit Used:
 50lb, 300, 301, 302, 303, 304, 305, 306, 307, 308, 309

Traceability:
 VA-05-3655G Cal Date: 02/11/2005 Recal Date: 02/11/2007 VA-05-3739 Cal Date: 02/11/2005 Recal Date: 02/11/2005

Best Uncertainties represent expanded uncertainties using a coverage factor of k=2 which provides a confidence of approximately 95%. Documented results relate only to the listed item tested. Weights used to calibrate this scale are traceable to NIST and inspected prior to use. Client accepts the judgement of Security Scale Service, Inc. in their procedure/test standard used for calibration unless otherwise indicated. This calibration report may not be reproduced without express written consent of Security Scale Service Inc. Security Scale welcomes all comments, any feedback or suggestions.

Authorized Distributor For:

- Avery Weigh-Tronix
- Fairbanks Scales
- Emery Winslow Scale Co.
- Rice Lake Weighing Systems

Accredited By:

American Association for Laboratory Accreditation



Calibration
ISO/IEC 17025



Security Scale Service, Inc.

1519 11th Street NE | Roanoke, VA 24012

Phone 800.572.2140 | Fax 540.362.5897

www.securityscale.com

- Bristol: 276-466-3500
- Harrisonburg: 540-434-2225
- Lynchburg: 434-237-0111
- Richmond: 804-264-1587
- Roanoke: 540-362-5800
- Suffolk: 757-523-2226



Security Scale Service, Inc.

Introduces



CRS

Certificate Retrieval System

Free online calibration program puts calibration data at your fingertips. Call today for a demo of CRS.

Security Scale Service, Inc. is a veteran owned company.



Security Scale Service, Inc. introduces the latest in calibration reporting technology. With the Certificate Retrieval System (CRS), your calibration data can be accessed online, from anywhere, at any time.

CRS allows you to view calibration information at a glance, including:

- Equipment status
- Location
- Adjustments made to equipment during inspection
- History of equipment
- Traceability (calibration report for weights, weight kits used in scale inspections)

How it Works

Each customer is set up on our secure website with an individual user name and password of choice. Our courteous Service Department staff assist in setup and will email account confirmation when the setup is complete. Instructions detailing the use of CRS are also sent via email.



Calibration Data Transfer

Our certified technicians perform a scale inspection and complete initial calibration reports on paper.

Calibration data is entered into a PDA, from which it is uploaded to the CRS system on our secure website.

Once the calibration data is uploaded to our website, it may be accessed at any date and time with the account user name and password.

Benefits of CRS

- Instant retrieval of calibration data
- Unlimited, immediate access to NIST certificates and most recent calibration reports
- Print multiple inspection reports sorted by date, scale, company or data category

Information Available through CRS

Scale Information

Scale ID	Serial Number	Location	Manufacturer	Scale Range	Status	Last Cert Date	Next Due Date
155	355-A	test	A/N/D 4323	0-5000 x 1 lb	Out of Service	05/12/2008	11/30/2008
124	23422	plant	WEIGHTRONIX 322-T100	0-10000 x 1 lb	In Service	06/18/2008	09/18/2008
109	1068	WAREHOUSE	WEIGHTRONIX W1-130	0-15 x 1 lb	In Service	05/25/2007	06/25/2007
162-4	362424	PRODUCTION	WEIGHTRONIX PC-B02	0-50 x 1 lb	In Service	05/25/2007	06/25/2007
112	0123	PRODUCTION	FAIRBANKS 413132	0-500 x 1 lb	In Service	05/15/2007	06/15/2007
123-1	123321	LAB	SARTORIUS CP42025	0.00-4200.00 x 0.01 g	In Service	05/15/2007	06/25/2007
156-4	456654	LAB	MOCONAB3	0.0-2.0 x 0.1 g	In Service	04/28/2007	06/28/2007
133	333-A	TRUCK SCALE	TRINER SCALE 755-57	0-80000 x 20 lb	In Service		
165-1	7654321	INBOUND TRUCK SCALE	WEIGHTRONIX W1125 LED	0-100000 x 20 lb	In Service		

Adjustments

Certificate ID	Date	Preload	Final	Within Accuracy	P/F/A	Technician
67890555-A0805131	05/12/2008	F	F	P	P	AARON JAMERSON
67890555-A080513	05/13/2008	P	P	P	P	AARON JAMERSON
67890555-A071127	11/27/2007	P	P	P	P	Mr. tech
67890555-A070525	05/24/2007	P	P	P	P	Mr. tech
67890458654070525	04/28/2007	P	P	P	P	Mr. tech
67890352424070525	05/25/2007	P	P	P	P	Mr. tech
6789022422080518	06/18/2008	P	P	P	P	Mr. tech
6789022422071127	11/27/2007	P	P	P	P	Mr. tech
6789022422070925	05/25/2007	P	P	P	P	Mr. tech
67890123321070525	05/15/2007	P	P	P	P	Mr. tech
678901098070525	05/25/2007	P	P	P	P	Mr. tech
6789023432070525	05/15/2007	P	P	P	P	Mr. tech

History

Certificate ID	Date	Preload	Final	Within Accuracy	P/F/A	Technician
67890555-A060619	02/28/2004	P	P	P	F	Mr. tech
01051003	10/03/2005	F	P	P	P	Mr. tech
67890555-A060509	11/17/2005	P	P	P	P	dustin
0100511182	11/18/2005	P	P	P	P	Mr. tech
67890010060213	02/13/2006	P	P	P	P	Mr. tech
67890010060302	03/02/2006	P	P	F	P	dbryan
67890555-A061110	11/10/2006	P	P	P	P	Mr. tech
67890555-A070326	03/26/2007	P	P	F	P	Mr. tech
67890555-A070525	05/24/2007	P	P	P	P	Mr. tech

Traceability

NIST Number	Calibration Date	Re-Calibration Date	Certificate
VA-07-52008	06/07/2007	06/07/2008	Download
NC0711-040-W	11/09/2007	11/08/2008	Download
VA-08-5557A	01/09/2008	01/09/2009	Download
VA-08-54557A	10/19/2007	10/19/2008	Download
4083-1751780 WEHREND	11/14/2007	11/13/2008	Download
APLA (SBB) SCOPE	08/28/2006	07/31/2008	Download
VA-08-5316A	08/13/2007	08/13/2008	Download
NC0803-005-W	03/03/2008	03/03/2010	Download